



The GS Testing Module for function tests of ECG devices

- ECG impulse output
- periodic signals
- arrhythmia output
- pacemaker outputs
- simulation of respiration
- signal output according to IEC 60601

Technical Data

ECG impulse amplitude :	1–5 mV ± 1% in 1 mV steps	ECG impulse frequency :	10 - 300 bpm ± 2 % in 1 bpm steps
ECG signal duration :	1–200 ms ± 1% in 1 ms steps	Signal frequency variable :	1 - 100 Hz ± 2 % in 1 Hz steps
ECG impulse form :	sinus sinus square triangle rectangle trapeze ISO ventricular fibrillation ventricular tachycardia mains frequency NSR	Signal frequency sine :	0,3 Hz ± 2 %
		Respiration :	Basic value ± 1 Ohm
		Accessories :	10 x STA8 ECG Adapter-Clip 1 x STA30 ECG plug adapter 1 x STA32 ECG adapter line

The GS test module ES serves as test-signal generator for ECG impulses. These ECG impulses can be used for the functional test of the signal-representation and signal-evaluation of ECG monitors. Furthermore, the extended ECG functionality can be tested over the respiration-function with apnoea-alarm-control.

The GS-X basic unit V4 or V5 and a PC are required for testing with this plug-in module.

The execution of a signal-output can be done with the PC-Software. The results of the signal-representation can be determined, assessed and stored.

PC software can be used for a flexible output of the offered waveforms. Consequently, an integration into complex test instructions and into automatic test-sequences is possible. For that reason, a high-quality documentation of the signal-representation and signal-evaluation of an ECG appliance is possible.

Because of the, as far as possible, freely configurable and variable ECG signal forms, a large variety of signal sequences can be brought to the output. The simulation of arrhythmic signal forms offers the possibility to test more complex devices in accordance with a standard.

The generation of the calibration impulse CAL after IEC 60601-2-51 enables practical operating function-controls for manufacturers and examiners.

-	Voltage
Abl. I (L - R)	+ 1,00 mV
Abl. II (F - R)	+ 1,56 mV
Abl. III (F - L)	- 0,56 mV
N - R	+ 0,59 mV
N - L	+ 1,59 mV
N - F	+ 2,15 mV
N - C1	+ 0,59 mV
N - C2	+ 1,18 mV
N - C3	+ 1,75 mV
N - C4	+ 2,03 mV
N - C5	+ 2,83 mV
N - C6	+ 3,35 mV

(Technical modifications and errors reserved. 09/2022)
